

Screwdrivers

Instructions for use

Omnigrip™ Mini



Omnigrip™



Unigrip™



Multi-unit



Cover Screw Brånemark® System Hexagon



Important – Disclaimer of Liability:

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, they are his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof.

Please note that some products detailed in this Instruction for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description:

Screwdrivers are reusable instruments which are used in conjunction with Nobel Biocare clinical screws, abutment screws, cover screws, prosthetic screws, prosthetic components (e.g. laboratory screws, abutments, healing abutments, impression copings), rescue tools (bone mill guides, abutment retrieval tool), and drill stops.

The “machine” versions of the screwdrivers feature a fitting compatible to ISO 1797-1 in order to connect the driver (using a wrench adapter) to the Manual Torque Wrench Prosthetic, while the “manual” versions of the screwdrivers have an attached handle to hold and turn the driver by hand. See Nobel Biocare Instructions for Use (IFU) IFU1098 for information regarding the Manual Torque Wrench Prosthetic. This IFU is available for download at ifu.nobelbiocare.com.

Table 1: Compatible Interfaces, Connections, and Instruments

Screwdriver	Screw Interface	Handpiece connection	Instruments
Omnigrip Screwdriver Manual	Omnigrip	n/a	n/a
Omnigrip Screwdriver Machine	Omnigrip	Instruments with ISO 1797 fitting	– Handle for Machine Instruments – Torque Wrench adapters
Omnigrip Mini Screwdriver Manual	Omnigrip Mini	n/a	– Impression Coping TCC
Omnigrip Mini Screwdriver Machine	Omnigrip Mini	Instruments with ISO 1797 fitting	– Handle for Machine Instruments – Torque wrench adapters – Impression Coping TCC

Unigrip Screwdriver Manual	Unigrip	n/a	– Abutment Retrieval Instrument Titanium CC – Drill stops – Impression Copings (CC, Tri-channel, External-hex, Multi-unit, Brånemark System, Zygoma) – Healing Abutments (CC, NobelReplace, Brånemark System, Zygoma) – Healing Cap Multi-unit – Slim Healing Abutments CC – Slim Temporary Abutment CC – Bone Mill Guide CC
Unigrip Screwdriver Machine	Unigrip	Instruments with ISO 1797 fitting	– Handle for Machine Instruments – Torque wrench adapters – Abutment Retrieval Instrument Titanium CC – Drill Stops – Guided Implant Mounts – Impression Copings (CC, Tri-channel, External-hex, Multi-unit, Brånemark System, Zygoma) – Healing Abutments (CC, NobelReplace, Brånemark System, Zygoma) – Healing Cap Multi-unit – Slim Healing Abutment CC – Slim Temporary Abutment CC
Screwdriver Manual Multi-unit	Internal Hex	n/a	– Abutment retrieval tool Nobel Biocare N1 TCC
Screwdriver Machine Multi-unit	Internal Hex	Instruments with ISO 1797 fitting	– Handle for Machine Instruments – Torque wrench adapters – Abutment retrieval tool Nobel Biocare N1 TCC
Screwdriver Manual M-u Bmk Syst WP	Internal Hex	n/a	n/a
Screwdriver Machine M-u Bmk Syst WP	Internal Hex	Instruments with ISO 1797 fitting	– Handle for Machine Instruments – Torque wrench adapters
Cover Screw Driver Bmk Syst Hexagon	External Hex	n/a	n/a

Screwdrivers Omnigrip™ and Omnigrip™ Mini (Manual and Machine):

The Screwdrivers Omnigrip™ and Omnigrip™ Mini are used to tighten screws and prosthetic components where the interface allows an angulation between screw and screwdriver of up to 25°. These screwdrivers can be used to engage the respective screw or component in order to pick up and transfer it from outside the oral cavity to the implant site, and to subsequently loosen or tighten the screw or component.

The Screwdrivers Omnigrip™ and Omnigrip™ Mini are available in both manual and machine versions and in various lengths and are compatible with screws and prosthetic components which feature the Omnigrip™ or Omnigrip™ Mini interface, respectively.

Screwdrivers Unigrip™ (Manual and Machine):

The Screwdrivers Unigrip™ are available in both manual and machine versions and in various lengths and are used to tighten and loosen screws, prosthetic components and abutment retrieval instruments which feature the Unigrip™ interface.

Screwdrivers Multi-unit (Manual and Machine) and Screwdriver Multi-unit Brånemark System WP (Machine and Manual):

The Screwdrivers Multi-unit are available in both manual and machine versions and are used to tighten and loosen the screws of the Multi-unit Abutments, as well as screws for abutment retrieval instruments, that feature the external hex interface.

The Screwdriver Multi-unit Brånemark System WP are available in both manual and machine versions and feature a larger external hex interface which is used with Multi-unit Abutment Brånemark System WP.

Screwdriver Cover Screw Brånemark System® Hexagon:

The Screwdriver Cover Screw Brånemark System Hex is used to tighten and loosen Cover Screws Brånemark System.

Intended Use / Intended Purpose:

Screwdrivers Manual and Machine:

Intended for use to tighten and/or loosen screws used to connect dental implant system components.

Indications:

Same as Intended Use / Intended Purpose.

Contraindications:

It is contraindicated to use Screwdrivers in:

- Patients who are medically unfit for an oral surgical procedure.
- Patients who are contraindicated for treatment with Nobel Biocare implants or restorative components.
- Patients who are allergic or hypersensitive to stainless steel or titanium nitride (TiN).

For contraindications specific to the screws, abutment, or other components, refer to the Nobel Biocare Instructions for Use for the respective component:

Table 2: Instruction For Use of Components

Components	Instruction For Use references
Clinical Screw, Abutment Screw, Prosthetic	IFU1057
Healing Abutment	IFU1026 and IFU1094
Manual Torque Wrenches Surgical and Prosthetic	IFU1098
Impression coping	IFU1086
Bone Mills and Bone Mill Guides	IFU1089
Abutment Retrieval Instrumentation	IFU1096
Drill Stop Kits for Guided and Freehand	IFU1092
Nobel Guided Surgery Tooling	IFU2011

Cautions:

General:

Close cooperation between surgeon, restorative dentist and dental laboratory technician is essential for a successful implant treatment.

It is strongly recommended that the screwdrivers are used only with compatible Nobel Biocare instruments and/or components. Use of instruments or components that are not intended to be used in combination with the screwdrivers can lead to product failure, damage to tissue, or unsatisfactory esthetic results.

When using a new device/treatment method for the first time, working with a colleague who is experienced with the new device/treatment method may help avoid possible complications. Nobel Biocare has a global network of mentors available for this purpose.

Before Surgery:

All components, instruments and tooling used during surgical and laboratory procedure must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components.

At Surgery:

Care and maintenance of sterile instruments are crucial for a successful treatment. Sterilized instruments not only safeguard your patients and staff against infection but are also essential for the outcome of the total treatment.

Because of the small sizes of the devices, care must be taken that they are not swallowed or aspirated by the patient. It is appropriate to use specific supporting tools to prevent aspiration of loose parts (e.g. gauze, dental dam, or a throat shield).

Intended Users and Patient Groups:

The screwdrivers are to be used by dental health care professionals.

The screwdrivers are to be used in patients subject to dental implant treatment.

Clinical Benefits and Undesirable Side Effects:

Clinical Benefits Associated with Screwdrivers:

Screwdrivers are components of treatment with a dental implant system and/or dental crowns and bridges. As a clinical benefit of treatment, patients can expect to have their missing teeth replaced and/or crowns restored.

Undesirable Side Effects Associated with Screwdrivers:

The use of this device is part of an invasive treatment which may be associated with typical side effects such as inflammation, infection, bleeding, hematoma, pain, swelling. Depending on location it may also lead in rare cases to fenestration or fracture of bone, perforation of neighboring structures, sinusitis, or sensory/motor disturbances. During use of this device the pharyngeal (gag) reflex may be triggered in patients with a sensitive gag reflex.

Notice regarding serious incidents:

For a patient/user/third party in the European Union and in countries with an identical regulatory regime (Regulation 2017/745/EU on Medical Devices); if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and to your national authority. The contact information for the manufacturer of this device to report a serious incident is as follows:

Nobel Biocare AB

<https://www.nobelbiocare.com/complaint-form>

Handling Procedure:

Manual Screwdrivers:

Note: Prior to use, loop dental floss through the hole in the handle of the manual screwdriver to prevent dropping the instrument, potentially into the mouth of the patient where it may be aspirated or swallowed.

- Engage the screwdriver to the screw or component with light pressure.
- Tighten or loosen the screw/component by hand.

Machine Screwdrivers:

- Connect the screwdriver to the Manual Torque Wrench Adapter Prosthetic.
- Engage the screwdriver to the screw or component with light pressure.
- Connect the Manual Torque Wrench Prosthetic to the screwdriver/wrench adapter assembly and tighten the screw/component to the recommended tightening torque. For the maximum tightening torque of the screws or components which are compatible with the screwdrivers refer to the IFU for the screw/component. The maximum allowed tightening torque of the screwdrivers is presented in **Table 3**.

Table 3: Maximum Tightening Torque for Machine and Manual Screwdrivers

Screwdriver	Maximum Tightening Torque
Omnigrip™	35 Ncm
Omnigrip™ Mini	20 Ncm
Unigrip™	35 Ncm
Multi-unit	35 Ncm
Cover Screw Brånemark® System Hexagon	Hand-tighten

Caution: Never exceed recommended maximum tightening torque in applicable instructions for use of the surgical or prosthetic screw/component. Overtightening of the screw may lead to a screw fracture and/or damage of the component.

Caution: In case the Screwdriver Omnigrip™ or Omnigrip™ Mini is used at an angulation to the screw and slips out of the interface, increase the axial force applied on the screwdriver, or try reducing the angulation of the screwdriver to the screw.

Note: Machine screwdrivers can be connected to the Handle for Machine Instruments instead of the torque wrench, and can then be used manually. Refer to Nobel Biocare IFU1090 for more information regarding the Handle for Machine Instruments.

Materials:

Screwdrivers Omnigrip™, Omnigrip™ Mini and Unigrip™ (Manual and Machine): Stainless steel AISI 303 / AISI 304 / 420F Mod according to ASTM F899, with titanium nitride (TiN) coating. Screwdrivers Multi-unit (Manual and Machine) and Cover Screw Brånemark System Hexagon: Stainless steel AISI 303 / AISI 304 / 420F Mod according to ASTM F899.

Sterility and Reusability Information:

Sterility and Reusability Claims for Product Category II:

Screwdrivers delivered non-sterile and are intended for reuse. Prior to use clean and sterilize the product following the manual or automated procedure in the Cleaning and Sterilization Instructions.

Warning: Use of non-sterile device may lead to infection of tissues or infectious diseases.

The screwdrivers are reusable instruments which shall be inspected before each re-use to ensure that the integrity and performance continues to be maintained. Check if any wear, abrasion of the coating, deformations or corrosion is visible on the instrument. Screwdrivers showing those signs shall be discarded.

If with slight pressure the Screwdriver Omnigrip and Omnigrip Mini do not engage in the respective screw, the screwdriver is worn and shall be discarded.

Warning: Do not use device if the packaging has been damaged or previously opened.

Note: Screwdrivers can be processed as individual devices as described in the Cleaning and Sterilization Instructions below, or together with other devices in a PureSet tray following the cleaning and sterilization instructions in Nobel Biocare Instructions for Use (IFU) IFU1067. This IFU is available on ifu.nobelbiocare.com.

Cleaning and Sterilization Instructions:

Screwdrivers are delivered non-sterile by Nobel Biocare and are intended for reuse. Prior to each use, the devices must be cleaned and sterilized by the user.

The devices can be cleaned manually, or in an automatic washer. Each device must then be individually sealed in a sterilization pouch and sterilized.

The following cleaning and sterilization processes have been validated according to international standards and guidelines as applicable:

- Manual and Automated Cleaning: AAMI TIR 12.
- Sterilization: AAMI ST79 and ISO 17665-1.

According to EN ISO 17664, it is the responsibility of the user/processor to ensure that the processing/reprocessing is performed using equipment, materials and personnel which are suitable to ensure the effectiveness of the processes. Any deviation from the following instructions should be validated by the user/processor to ensure the effectiveness of the process.

Note: The manufacturer's instructions for use for any detergent / cleaning solution and/or equipment and accessories used to clean and/or dry the device(s) must be strictly followed where applicable.

Note: The Screwdrivers have been validated to withstand these cleaning and sterilization procedures.

Caution: Do not deviate from the following reprocessing instructions.

Caution: Keep dissimilar metals separated during sterilization to resist corrosion.

Initial Treatment at Point of Use Prior to Reprocessing:

- Discard single-use instruments and worn reusable instruments immediately after use.
- Remove excess soil and debris from reusable devices to be reprocessed using absorbent paper wipes. Use a dental probe to remove soil and debris from cavities, where applicable.
- Rinse the devices with cold running tap water.

Containment and Transportation/Shipping to Reprocessing Area:

- After removal of excess soil and debris, store the devices in a container which is suitable to protect the devices during transportation and to avoid any contamination of personnel or the environment.
- Transport the devices to the reprocessing area as soon as practical. If transfer to the processing area is likely to be delayed, consider covering the devices with a damp cloth or store it in a closed container to avoid drying of soil and/or debris.

Note: Reusable devices should be reprocessed by initiating the prescribed automated or manual cleaning and drying procedures within 1 hour of use, to ensure the efficacy of the reprocessing.

- If the devices are shipped to an outside facility for reprocessing, they must be contained in a transportation or shipping container which is suitable to protect the devices during transportation and to prevent contamination of personnel or the environment.

Automated Cleaning and Drying (Including Pre-cleaning):

Pre-cleaning:

- Immerse the device in 0.5% lukewarm enzymatic cleaning agent (e.g. Neodisher Medizym) for a minimum of 5 minutes.
- Fill lumina (where applicable) with 0.5% lukewarm enzymatic cleaning agent (e.g. Neodisher Medizym) using a 20 ml syringe.

- Brush the outer surfaces with a soft bristled nylon brush (e.g. Medsafe MED-100.33) for a minimum of 20 seconds until all visible soil is removed.
- Brush the inner surfaces, lumina and cavities (where applicable) with an appropriately sized bottle brush (e.g. 1.2 mm / 2.0 mm / 5.0 mm diameter) for a minimum of 20 seconds until all visible soil is removed.
- Thoroughly rinse all outer and inner surfaces, lumina and cavities (where applicable) with cold running tap water for a minimum of 10 seconds to remove all cleaning solution.
- Rinse lumina (where applicable) with 20 ml tap water using a 20 ml syringe.

Automated Cleaning and Drying:

The following washer was used in the Nobel Biocare validation: Miele G7836 CD with the Vario TD program.

Note: It is recommended to perform the automated cleaning and drying with a maximum load of 11 individual devices.

- Place the devices in a suitable rack or load carrier (e.g. metal sieve basket).
- Load the devices into the washer. Ensure the rack or load carrier is oriented in a horizontal position.
- Perform automatic cleaning. The following parameters are based on the Vario TD program on the Miele G7836 CD washer:
 - Minimum 2 minutes pre-cleaning with cold tap water.
 - Draining.
 - Minimum 5 minutes cleaning with minimum 55°C (131°F) tap water and 0.5% mildly alkaline detergent (e.g. Neodisher Mediclean).
 - Draining.
 - Minimum 3 minutes neutralization with cold desalinated water.
 - Draining.
 - Minimum 2 minutes rinsing with cold desalinated water.
 - Draining.
- Run drying cycle at minimum 50°C (122°F) for a minimum of 10 minutes.
- Dry with compressed air or clean and lint-free single use wipes, if any residual moisture remains after the drying cycle.

Visual Inspection:

After cleaning and drying, inspect the device for unacceptable deterioration such as corrosion, discoloration, pitting, or cracked seals and properly discard any devices that fail the inspection.

Manual Cleaning and Drying:

- Immerse device for a minimum of 5 minutes in a sterile 0.9% NaCl solution.
- Scrub the outer surfaces of the device with soft-bristled nylon brush for a minimum of 20 seconds until all visible soil is removed.
- Flush the inner surfaces, lumina and cavities (where applicable) with 20 ml lukewarm enzymatic cleaning solution (e.g. Cydezyme ASP; maximum 45°C (113°F)) using an irrigation needle connected to a 20 ml syringe.
- Brush the inner surfaces, lumina and cavities (where applicable) with appropriately sized bottle brush (e.g. 1.2 mm / 2.0 mm / 5.0 mm diameter) for a minimum of 10 seconds until all visible soil is removed.
- Thoroughly rinse the outer surfaces and lumina of the device with cold running tap water for a minimum of 10 seconds to remove all cleaning solution.
- Immerse the device in an ultrasonic bath (e.g. Bandelin; frequency 35 kHz; effective ultrasonic power 300 W) containing 0.5% enzymatic cleaning agent (e.g. Cydezyme ASP) and treat for a minimum of 5 minutes at minimum 40°C (104°F) / maximum 45°C (113°F).
- Flush the inner surfaces, lumina and cavities (where applicable) with 20 ml lukewarm tap water using an irrigation needle connected to a 20 ml syringe.
- Thoroughly rinse the outer surfaces of the device with purified or sterile water for a minimum of 10 seconds to remove all cleaning agent.
- Dry with compressed air or clean and lint-free single use wipes.

Visual Inspection:

After cleaning and drying, inspect the device for unacceptable deterioration such as corrosion, discoloration, pitting, cracked seals and properly discard any devices that fail the inspection.

Sterilization:

The following steam sterilizers were used in the Nobel Biocare validation: Systec HX-320 (pre-vacuum cycle), Amsco Century Sterilizer (gravity cycle).

Note: It is recommended to perform sterilization with a maximum load of 11 devices individually sealed in sterilization pouches.

- Seal each device in a suitable sterilization pouch. The sterilization pouch should fulfill the following requirements:
 - EN ISO 11607 and/or DIN 58953-7.

- Suitable for steam sterilization (temperature resistance up to at least 137°C (279°F), sufficient steam permeability).
- Sufficient protection of the instruments as well as of the sterilization packaging to mechanical damage.

Table 4 presents examples of suitable sterilization containers, pouches, and wraps.

Table 4: Recommended Sterilization Pouches

Method	Recommended Sterilization Pouch
Gravity Cycle	SPSmedical Self-Seal sterilization pouch
Pre-vacuum Cycle	SteriCLIN® pouch

- Label the sterilization pouch with information necessary to identify the device (for example, the product name with article number and lot/batch number (if applicable)).
- Place the sealed sterilization pouch into the autoclave/sterilizer. Ensure that the sterilization pouch is oriented in a horizontal position.
- Sterilize the device. Both the gravity displacement cycle and pre-vacuum (top dynamic air removal) cycle can be applied, using the following recommended parameters (Table 5):

Table 5: Recommended Sterilization Cycles

Cycle	Minimum Temperature	Minimum Sterilization Time	Minimum Drying Time (In Chamber)	Minimum Pressure
Gravity Cycle ¹	132°C (270°F)	15 minutes	20 minutes	≥2868.2 mbar ⁴
Pre-Vacuum Cycle ¹	132°C (270°F)	4 minutes		
Pre-Vacuum Cycle ²	134°C (273°F)	3 minutes		≥3042 mbar ⁵
Pre-Vacuum Cycle ³	134°C (273°F)	18 minutes		

¹ Validated sterilization processes to achieve a Sterility Assurance Level (SAL) of 10⁻⁶ in accordance to EN ISO 17665-1.

² Recommendation of the Welsh Health Technical Memorandum (WHTM) 01-01 Part C.

³ Recommendation of the World Health Organization (WHO) for steam sterilization of instruments with potential TSE/CJD contamination. Ensure that the packaging and monitoring systems (chemical/biological indicators) used for this cycle are validated for these conditions.

⁴ Saturated steam pressure at 132°C as per required by EN ISO 17665-2.

⁵ Saturated steam pressure at 134°C as per required by EN ISO 17665-2.

Note: Autoclave/sterilizer design and performance can affect the efficacy of the sterilization process. Healthcare facilities should therefore validate the processes that they use, employing the actual equipment and operators that routinely process the devices. All autoclaves/sterilizers should comply with the requirements of, and be validated, maintained and checked in accordance to SN EN 13060, EN 285, EN ISO 17665-1, and/or AAMI ST79, or to the applicable national standard. The autoclave/sterilizer manufacturer's instructions for use must be strictly followed.

Storage and Maintenance:

After sterilization, place the labeled and sealed sterilization pouch in a dry and dark place. Follow the instructions provided by the manufacturer of the sterilization pouch regarding the storage conditions and expiration date of the sterilized device.

Containment and Transportation/Shipping to Point of Use:

The container and/or outer packaging used to transport or ship the reprocessed device back to the point of use must be suitable to protect and safeguard the sterility of the devices during transportation, taking the device packaging and the required transportation or shipping process (intrafacility transportation or shipping to an external site) into account.

Performance Requirements and Limitations:

To achieve the desired performance, the screwdrivers must only be used with the products described in this Instructions for Use and/or in the Instructions for Use for other compatible Nobel Biocare products, and in accordance with the Intended Use for each product. To confirm the compatibility of products which are intended to be used in conjunction with the screwdrivers, check the color coding, dimensions, lengths, connection type and/or any direct marking as applicable on the products or product labeling.

Facilities and Training:

It is strongly recommended that clinicians, new as well as experienced implant users, always go through special training before undertaking a new treatment method. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more info please visit www.nobelbiocare.com.

Storage, Handling and Transportation:


The device must be stored and transported in dry conditions in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage and transportation may influence device characteristics leading to failure.

Disposal:

Safely discard potentially contaminated or no longer usable medical devices as healthcare (clinical) waste in accordance with local healthcare guidelines, country and government legislation or policy.

Separation, re-cycling or disposal of packaging material shall follow local country and government legislation on packaging and packaging waste, where applicable.

Manufacturer and Distributor Information:

Manufacturer:
 Nobel Biocare AB
 Box 5190, 402 26
 Västra Hamngatan 1
 411 17 Göteborg
 Sweden
www.nobelbiocare.com

Distributed in Australia by:
 Nobel Biocare Australia Pty Ltd
 Level 4/7 Eden Park Drive
 Macquarie Park, NSW 2114 Australia
 Phone: +61 1800 804 597

Distributed in New Zealand by:
 Nobel Biocare New Zealand Ltd
 33 Spartan Road
 Takanini, Auckland, 2105 New Zealand
 Phone: +64 0800 441 657



Note: Refer to the product label to determine the applicable CE mark for each device.

Basic UDI-DI Information:

The following table lists the Basic UDI-DI information for the devices described in this IFU.

Product	Basic UDI-DI Number
Omnigrip™ Screwdrivers Manual	73327470000001787L
Screwdrivers Manual Omnigrip™ Mini	73327470000001787L
Screwdrivers Manual Unigrip™	73327470000001787L
Screwdrivers Manual Multi-unit	73327470000001787L
Screwdriver Manual Multi-unit Brånemark System WP	73327470000001787L
Cover Screw Driver Brånemark System Hexagon	73327470000001787L
Omnigrip™ Screwdrivers Machine	73327470000001797N
Screwdrivers Machine Omnigrip™ Mini	73327470000001797N
Screwdrivers Machine Unigrip™	73327470000001797N
Screwdriver Machine Multi-unit	73327470000001797N
Screwdriver Machine Multi-unit Brånemark System WP	73327470000001797N

Symbols Glossary:

The following symbols may be present on the device labeling or in information accompanying the device. Refer to the device labeling or accompanying information for the applicable symbols.



Authorized Representative in the European Union



Batch code



Catalogue number



Caution



CE mark



CE mark with Notified Body number



Consult instructions for use



Contains hazardous substances



Implantation date



Date of manufacture



Do not re-sterilize



Do not re-use



Do not use if package is damaged



Double sterile barrier system



For prescription use only



Health care centre or doctor



Keep away from sunlight



Keep dry

symbol.glossary.nobelbiocare.com/ifu.nobelbiocare.com

Link to Online Symbols Glossary and IFU Portal



Magnetic resonance conditional



Manufacturer



Medical device



Non-pyrogenic



Non-sterile



Patient identification



Patient information website



Patient number



Contains or presence of DEHP phthalate



Contains or presence of phthalate



Contains or presence of natural rubber latex



Magnetic resonance safe



Serial number



Single sterile barrier system



Single sterile barrier system with protective packaging inside



Single sterile barrier system with protective packaging outside



Sterilized using ethylene oxide



Sterilized using irradiation



Temperature limit



Tooth number



Upper limit of temperature



Sterilized using steam or dry heat



Unique Device Identifier



Use-by date

EN All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images in this folder are not necessarily to scale. All product images are for illustration purposes only and may not be an exact representation of the product.