

Abutment Retrieval Instrumentation

Instructions for use



Important – Disclaimer of Liability:

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, they are his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof. Please note that some products detailed in this Instruction for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description:

Abutment retrieval instruments are used to remove dental implant abutments or finalized prosthetic restorations that have become stuck to a dental implant, after the abutment screw or clinical screw attaching the abutment to the implant has been removed. There are two types of abutment retrieval instruments, one type for removing zirconia abutments, and the other for removing titanium abutments.

Abutment Retrieval Instrument Zirconia CC:

The Abutment Retrieval Instrument Zirconia CC is used to remove zirconia abutments. They are comprised of two pieces: one is a hollow cylinder ("engaging pin", Figure A) that is placed through the screw access hole of the zirconia abutment/prosthesis, and the second piece (Figure B) is the "activating pin" that is inserted through the engaging pin. After using forceps to compress the two components, the engaging pin engages the abutment and lifts it vertically, so that the abutment can be removed by hand.



Figures A and B: Hollow Cylinder (A) and Activating Pin (B) of Two-piece Abutment Retrieval Instrument Zirconia CC

The Abutment Retrieval Instrument Zirconia CC is available in the NP and RP/WP platform sizes and is compatible with Nobel Biocare zirconia abutments (see Table 1).

Table 1: Abutments compatible with Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Instrument Zirconia CC

	Abutment Description	Retrieval Tool	Screwdriver
NP	Temporary Abutment Engaging NP	Abutment Retrieval Instrument Titanium CC NP	Screwdriver Unigrip
	Esthetic Abutment NP		
	Snappy™ Abutment NP		
	NobelProcera® Abutment Titanium NP		
RP/WP	Temporary Abutment Engaging RP/WP	Abutment Retrieval Instrument Titanium CC RP/WP	Screwdriver Unigrip
	Esthetic Abutment RP/WP		
	Snappy™ Abutment RP/WP		
	NobelProcera® Abutment Titanium RP/WP		
NP	NobelProcera® Abutment Zirconia NP	Abutment Retrieval Instrument Zirconia CC NP	Forceps
	Adapter for Zirconia Abut. CC NP (NobelProcera® ASC Abutment Zirconia NP, NobelProcera® FCZ Implant Crown NP)		
RP/WP	NobelProcera® Abutment Zirconia RP	Abutment Retrieval Instrument Zirconia CC RP/WP	Forceps
	Adapter for Zirconia Abut. CC RP and WP (NobelProcera® ASC Abutment Zirconia RP/WP, NobelProcera® FCZ Implant Crown RP/WP)		

Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Tool Nobel Biocare N1™ TCC:

The Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Tool Nobel Biocare N1™ TCC are used to remove titanium abutments. They consist of a pin with a threaded portion that is threaded through the abutment inner threads (see Figure C). By applying torque with the screwdriver, the unthreaded portion of the pin comes into contact with the implant, which pushes the abutment up so that it can be removed by hand.



Figure C: Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Tool Nobel Biocare N1™ TCC

The Abutment Retrieval Instrument Titanium CC is available in the NP (magenta) and RP/WP (silver) platform sizes and is compatible with Nobel Biocare titanium abutments (see Table 1).

The Abutment Retrieval Tool Nobel Biocare N1™ TCC is available in the NP (magenta) and RP (silver) platform sizes and is compatible with Nobel Biocare titanium abutments (see Table 2).

Table 2: Abutments compatible with Abutment Retrieval Tool Nobel Biocare N1™ TCC

	Abutment Description	Retrieval Tool	Screwdriver
NP	Healing Abutment Nobel Biocare N1™ TCC NP	Abutment Retrieval Tool Nobel Biocare N1™ TCC NP	Screwdriver Machine/ Manual Multi-unit
	Temporary Abutment Nobel Biocare N1™ TCC NP*		
	Universal Abutment Nobel Biocare N1™ TCC NP**		
	Multi-unit Abutment Xeal™ Nobel Biocare N1™ TCC NP		
	17° Multi-unit Abut Xeal™ Nobel Biocare N1™ TCC NP		
RP	Healing Abutment Nobel Biocare N1™ TCC RP	Abutment Retrieval Tool Nobel Biocare N1™ TCC RP	Screwdriver Machine Multi-unit
	Temporary Abutment Nobel Biocare N1™ TCC RP*		
	Universal Abutment Nobel Biocare N1™ TCC RP**		
	Multi-unit Abutment Xeal™ Nobel Biocare N1™ TCC RP		
	17° Multi-unit Abut Xeal™ Nobel Biocare N1™ TCC RP		
	30° Multi-unit Abut Xeal™ Nobel Biocare N1™ TCC RP		

* Abutment Retrieval Tool only compatible if post length (upper part of the abutment) is reduced below 50%. Alternatively, use grasping instruments, e.g. forceps.

** Widening of crown screw access hole is required to fit the Abutment Retrieval Tool into the abutment.

Intended Use/Intended Purpose:

Abutment Retrieval Instrument Zirconia CC, Abutment Retrieval Instrument Titanium CC, Abutment Retrieval Tool Nobel Biocare N1™ TCC:
Intended for use to facilitate the removal of dental implant system components.

Indications:

Abutment Retrieval Instrument Zirconia CC:

The Abutment Retrieval Instrument Zirconia are indicated for use to facilitate the removal of zirconia abutments from an endosseous dental implant placed in the maxilla or mandible.

Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Tool Nobel Biocare N1™ TCC:

The Abutment Retrieval Instrument Titanium CC and Abutment Retrieval Tool Nobel Biocare N1™ TCC are indicated for use to facilitate the removal of titanium abutments from an endosseous dental implant placed in the maxilla or mandible.

Contraindications:

In general, contraindications are applicable for implant surgery related procedures in:

- Patients who are medically unfit for an oral surgical procedure.
- Patients who are allergic or hypersensitive to medical grade stainless steel, stainless steel or commercially titanium alloy grade 5 (Ti6Al4V).

Cautions:

General:

It is strongly recommended that the abutment retrieval instruments are used only with compatible Nobel Biocare abutments (see Table 1 and Table 2). Use of abutments that are not intended to be used in combination with the abutment retrieval instruments can lead to product failure, damage to tissue, or unsatisfactory esthetic results.

When using a new device/treatment method for the first time, working with a colleague who is experienced with the new device/treatment method may help avoid possible complications. Nobel Biocare has a global network of mentors available for this purpose.

Before Surgery:

The device has not been evaluated in pediatric/adolescent patients and is not recommended for use in children. Routine treatment is not recommended until the end of the juvenile jaw bone growth phase has been properly documented.

Pre-operative hard tissue or soft tissue deficits may yield a compromised esthetic result or unfavorable implant angulations.

All components, instruments and tooling used during the clinical or laboratory procedure must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components.

At Surgery:

Care and maintenance of sterile instruments are crucial for a successful treatment. Sterilized instruments not only safeguard your patients and staff against infection but are also essential for the outcome of the total treatment.

Because of the small size of the devices, care must be taken that they are not swallowed or aspirated by the patient. It is appropriate to use specific supporting tools to prevent aspiration of loose parts (e.g. a throat shield).

Intended Users and Patient Groups:

Abutment Retrieval Instruments are to be used by dental health care professionals.

Abutment Retrieval Instruments are to be used in patients subject to dental implant treatment.

Clinical Benefits and Undesirable Side Effects:

Clinical Benefits Associated with Abutment Retrieval Instruments:

Abutment Retrieval Instruments are a component of treatment with a dental implant system and/or dental crowns and bridges. As a clinical benefit of treatment, patients can expect to have their missing teeth replaced and/or crowns restored.

Undesirable Side Effects Associated with Abutment Retrieval Instruments:

The use of this device is part of an invasive treatment which may be associated with typical side effects such as inflammation, infection, bleeding, hematoma, pain, swelling. Depending on location it may also lead in rare cases to fenestration or fracture of bone, perforation of neighboring structures, sinusitis, or sensory/motor disturbances. During use of this device the pharyngeal (gag) reflex may be triggered in patients with a sensitive gag reflex.

Notice regarding serious incidents:

For a patient/user/third party in the European Union and in countries with an identical regulatory regime (Regulation 2017/745/EU on Medical Devices); if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and to your national authority. The contact information for the manufacturer of this device to report a serious incident is as follows:

Nobel Biocare AB

<https://www.nobelbiocare.com/complaint-form>

Procedure for Retrieval of Zirconia Abutments Using the Abutment Retrieval Instrument Zirconia CC:

These tools are used to remove zirconia abutments when the abutment screw or clinical screw has been removed (see Figures A and B) using a screwdriver, but the abutment cannot be removed due to a tight conical seal.

1. The abutment/clinical screw must be unthreaded from both the internal threads of the implant and the abutment. In case the loose abutment/clinical screw is difficult to remove, use a small amount of sticky wax on the tip of the Screwdriver Unigrip™, which will aid in retention of the abutment screw head.
2. Insert the engaging pin (Figure A) into the abutment until it reaches a stop.

Note: The engaging pin has to be pushed in rather firmly to reach its end stop. There is first an intermediate stop that has to be passed before the pin is in final position.

3. Assemble the instrument by inserting the activating needle (Figure B).
4. Squeeze the abutment retrieval instrument parts together using e.g. a hemostat or pliers until the abutment is released (Figure D).

In case the Adapter for Zirconia Abutment has been used:

1. Remove the abutment so that only the adapter stays in place
2. Insert the engaging pin into the adapter until it reaches a stop

Note: The engaging pin has to be pushed in rather firmly to reach its end stop. There is first an intermediate stop that has to be passed before the pin is in final position.

3. Assemble the instrument by inserting the activating needle (Figure B).
4. Squeeze the abutment retrieval instrument parts together using e.g. a hemostat or pliers until the adapter is released (Figure D).



Figure D: Squeezing of abutment retrieval instrument parts together

Procedure for Retrieval of Titanium Abutments using Abutment Retrieval Instrument Titanium CC or Abutment Retrieval Tool Nobel Biocare N1™ TCC:

These tools are used to remove titanium abutments when the abutment screw or clinical screw has been removed but the abutment cannot be removed due to a tight conical seal.

Note: The abutment screw must be unthreaded from both the internal threads of the implant and the abutment. In case the loose abutment/clinical screw is difficult to remove, use a small amount of sticky wax on the tip of the screwdriver which will aid in retention of the abutment screw head.

1. Insert the Abutment Retrieval Instrument into the abutment and screw into place using the screwdriver until the tip of the screw touches the bottom of the hole inside the implant (Figure E).
2. Apply torque to the screwdriver to release the abutment from the implant.



Figure E: Insertion of Abutment Retrieval Instrument into abutment

Materials:

- Abutment Retrieval Instrument Zirconia CC, and Abutment Retrieval Tool Nobel Biocare N1™ TCC: Stainless steel according to ASTM A895/F899 and ISO 5832-1.
- Abutment Retrieval Instrument Titanium CC: Titanium alloy according to ASTM F136 and ISO 5832-3.

Sterility and Reusability Information:

The Abutment Retrieval Instruments are delivered non-sterile and intended for reuse. Prior to first use and reuse clean, disinfect and/or sterilize using the recommended parameters.

Warning: Use of non-sterile device may lead to infection of tissues or infectious diseases.

The Abutment Retrieval Instruments are re-usable instruments which shall be inspected before each re-use to ensure that the integrity and performance continues to be maintained. Inspect the devices for signs of degradation that may limit the useful life of the devices, such as:

- Compromised legibility of the lasermarking.
- Visible corrosion.
- Mechanical wear/damage.

The Abutment Retrieval Instruments shall be disposed if any of these signs of degradation are evident.

Cleaning and Sterilization Instructions:

The Abutment Retrieval Instruments are delivered non-sterile and intended for reuse. Prior to each use, the devices must be cleaned and sterilized by the user.

The devices can be cleaned manually, or in an automatic washer. Each device must then be individually sealed in a sterilization pouch and sterilized.

The following cleaning and sterilization processes have been validated according to international standards and guidelines as applicable:

- Manual and Automated Cleaning: AAMI TIR 12.
- Sterilization: AAMI ST79 and ISO 17665-1.

According to EN ISO 17664, it is the responsibility of the user/processor to ensure that the processing/reprocessing is performed using equipment, materials and personnel which are suitable to ensure the effectiveness of the processes. Any deviation from the following instructions should be validated by the user/processor to ensure the effectiveness of the process.

Note: The manufacturer's instructions for use for any detergent/cleaning solution and/or equipment and accessories used to clean and/or dry the device(s) must be strictly followed where applicable.

Note: The devices have been validated to withstand these cleaning and sterilization procedures.

Caution: Do not deviate from the described reprocessing instructions

Caution: Keep dissimilar metals separated during sterilization to resist corrosion

Initial Treatment at Point of Use Prior to Reprocessing:

1. Discard single-use instruments and worn reusable instruments immediately after use.
2. Remove excess soil and debris from reusable devices to be reprocessed using absorbent paper wipes.
3. Rinse the devices with cold running tap water.

Containment and Transportation/Shipping to Reprocessing Area:

1. After removal of excess soil and debris, store the devices in a container which is suitable to protect the devices during transportation and to avoid any contamination of personnel or the environment.
2. Transport the devices to the reprocessing area as soon as practical. If transfer to the processing area is likely to be delayed, consider covering the devices with a damp cloth or store it in a closed container to avoid drying of soil and/or debris.

Note: Reusable devices should be reprocessed by initiating the prescribed automated or manual cleaning and drying procedures within 1 hour of use, to ensure the efficacy of the reprocessing.

3. If the devices are shipped to an outside facility for reprocessing, they must be contained in a transportation or shipping container which is suitable to protect the devices during transportation and to prevent contamination of personnel or the environment.

Automated Cleaning and Drying (Including Pre-cleaning):

Pre-cleaning:

1. Immerse the devices in 0.5% lukewarm enzymatic cleaning agent (e.g. Neodisher Medizym) for a minimum of 5 minutes.
2. Fill lumina (where applicable) with 0.5% lukewarm enzymatic cleaning agent (e.g. Neodisher Medizym) using a 20 ml syringe.
3. Brush the outer surfaces with a soft bristled nylon brush (e.g. Medsafe MED-100.33) for a minimum of 20 seconds until all visible soil is removed.
4. Brush the inner surfaces, lumina and cavities (where applicable) with an appropriately sized bottle brush (e.g. 1.2 mm/2.0 mm/5.0 mm diameter) for a minimum of 20 seconds until all visible soil is removed.
5. Thoroughly rinse all outer and inner surfaces, lumina and cavities (where applicable) with cold running tap water for a minimum of 10 seconds to remove all cleaning solution.
6. Rinse lumina (where applicable) with 20 ml tap water using a 20 ml syringe.

Automated Cleaning and Drying:

The following washer was used in the Nobel Biocare validation: Miele G7836 CD with the Vario TD program.

Note: It is recommended to perform the automated cleaning and drying with a maximum load of 11 individual devices.

1. Disassemble components prior to cleaning (applicable only for Abutment Retrieval Instrument Zirconia CC).
2. Place the devices in a suitable rack or load carrier (e.g. metal sieve basket).
3. Load the devices into the washer. Ensure the rack or load carrier is oriented in a horizontal position.
4. Perform automatic cleaning. The following parameters are based on the Vario TD program on the Miele G7836 CD washer:
 - Minimum 2 minutes pre-cleaning with cold tap water.
 - Draining.
 - Minimum 5 minutes cleaning with minimum 55°C (131°F) tap water and 0.5% mildly alkaline detergent (e.g. Neodisher Mediclean).
 - Draining.
 - Minimum 3 minutes neutralization with cold desalinated water.
 - Draining.
 - Minimum 2 minutes rinsing with cold desalinated water.
 - Draining.
5. Run drying cycle at minimum 50°C (122°F) for a minimum of 10 minutes.
6. Dry with compressed air or clean and lint-free single use wipes, if any residual moisture remains after the drying cycle.

Visual Inspection:

After cleaning and drying, inspect the device for unacceptable deterioration such as corrosion, discoloration, pitting, or cracked seals and properly discard any devices that fail the inspection.

Manual Cleaning and Drying:

- Disassemble components prior to cleaning (applicable only for Abutment Retrieval Instrument Zirconia CC).
- Immerse devices for a minimum of 5 minutes in a sterile 0.9% NaCl solution.
- Scrub the outer surfaces of the device with soft-bristled nylon brush for a minimum of 20 seconds until all visible soil is removed.
- Flush the inner surfaces, lumina and cavities (where applicable) with 20 ml lukewarm enzymatic cleaning solution (e.g. Cidezyme ASP; maximum 45°C (113°F)) using an irrigation needle connected to a 20 ml syringe.
- Brush the inner surfaces, lumina and cavities (where applicable) with appropriately sized bottle brush (e.g. 1.2 mm/2.0 mm/5.0 mm diameter) for a minimum of 10 seconds until all visible soil is removed.
- Thoroughly rinse the outer surfaces and lumina of the device with cold running tap water for a minimum of 10 seconds to remove all cleaning solution.
- Immerse the device in an ultrasonic bath (e.g. Bandelin; frequency 35 kHz; effective ultrasonic power 300 W) containing 0.5% enzymatic cleaning agent (e.g. Cidezyme ASP) and treat for a minimum of 5 minutes at minimum 40°C (104°F)/maximum 45°C (113°F).
- Flush the inner surfaces, lumina and cavities of the device for a minimum of 15 seconds using a water jet pistol.
- Thoroughly rinse the outer surfaces of the device with purified or sterile water for a minimum of 10 seconds to remove all cleaning agent.
- Dry with compressed air or clean and lint-free single use wipes.

Visual Inspection:

After cleaning and drying, inspect the device for unacceptable deterioration such as corrosion, discoloration, pitting, cracked seals and properly dispose any devices that fail the inspection.

Sterilization:

The following steam sterilizers were used in the Nobel Biocare validation: Systec HX-320 (pre-vacuum cycle); Amsco Century Sterilizer (gravity cycle).

Note: It is recommended to perform sterilization with a maximum load of 11 devices individually sealed in sterilization pouches.

- Disassemble the components prior to cleaning (applicable only for Abutment Retrieval Instrument Zirconia CC).
- Seal each device in a suitable sterilization pouch. The sterilization pouch should fulfill the following requirements:
 - EN ISO 11607 and/or DIN 58953-7.
 - Suitable for steam sterilization (temperature resistance up to at least 137°C (279°F), sufficient steam permeability).
 - Sufficient protection of the instruments as well as of the sterilization packaging to mechanical damage.

Table 3 presents examples of suitable sterilization containers, pouches, and wraps.

Table 3: Recommended Sterilization Pouches

Method	Recommended Sterilization Pouch
Gravity Cycle	SPSmedical Self-Seal sterilization pouch
Pre-vacuum Cycle	SteriCLIN® pouch

- Label the sterilization pouch with information necessary to identify the device (for example, the product name with article number and lot/batch number (if applicable)).
- Place the sealed sterilization pouch into the autoclave/sterilizer. Ensure that the sterilization pouch is oriented in a horizontal position.
- Sterilize the device. Both the gravity displacement cycle and pre-vacuum (top dynamic air removal) cycle can be applied, using the following recommended parameters (Table 4):

Table 4: Recommended Sterilization Cycles

Cycle	Minimum Temperature	Minimum Sterilization Time	Minimum Drying Time (In Chamber)	Minimum Pressure
Gravity Cycle ¹	132°C (270°F)	15 minutes	20 minutes	≥2868.2 mbar ⁴
Pre-Vacuum Cycle ¹	132°C (270°F)	4 minutes		≥3042 mbar ⁵
Pre-Vacuum Cycle ²	134°C (273°F)	3 minutes		
Pre-Vacuum Cycle ³	134°C (273°F)	18 minutes		

¹ Validated sterilization processes to achieve a Sterility Assurance Level (SAL) of 10⁻⁶ in accordance to EN ISO 17665-1.

² Recommendation of the Welsh Health Technical Memorandum (WHTM) 01-01 Part C.

³ Recommendation of the World Health Organization (WHO) for steam sterilization of instruments with potential TSE/CJD contamination. Ensure that the packaging and monitoring systems (chemical/biological indicators) used for this cycle are validated for these conditions.

⁴ Saturated steam pressure at 132°C as per required by EN ISO 17665-2.

⁵ Saturated steam pressure at 134°C as per required by EN ISO 17665-2.

Note: Autoclave/sterilizer design and performance can affect the efficacy of the sterilization process. Healthcare facilities should therefore validate the processes that they use, employing the actual equipment and operators that routinely process the devices. All autoclaves/sterilizers should comply with the requirements of, and be validated, maintained and checked in accordance to SN EN 13060, EN 285, EN ISO 17665-1, and/or AAMI ST79, or to the applicable national standard. The autoclave/sterilizer manufacturer's instructions for use must be strictly followed.

Storage and Maintenance:

After sterilization, place the labeled and sealed sterilization pouch in a dry and dark place. Follow the instructions provided by the manufacturer of the sterilization pouch regarding the storage conditions and expiration date of the sterilized device.

Containment and Transportation/Shipping to Point of Use:

The container and/or outer packaging used to transport or ship the reprocessed device back to the point of use must be suitable to protect and safeguard the sterility of the devices during transportation, taking the device packaging and the required transportation or shipping process (interfacility transportation or shipping to an external site) into account.

Performance Requirements and Limitations:

To achieve the desired performance, the abutment screw retrieval instrumentation must only be used with the products described in this Instructions for Use and/or in the Instructions for Use for other compatible Nobel Biocare products, and in accordance with the Intended Use/Intended Purpose for each product. To confirm the compatibility of products which are intended to be used in conjunction with the abutment screw retrieval instrumentation, check the color coding, dimensions, lengths, connection type and/or any direct marking as applicable on the products or product labeling.

Facilities and Training:

It is strongly recommended that new and experienced users of Nobel Biocare products always go through special training before using a new product for the first time. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more information please visit www.nobelbiocare.com

Storage, Handling and Transportation:

The device must be stored and transported in dry conditions in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage and transportation may influence device characteristics leading to failure.

Disposal:

Safely discard potentially contaminated or no longer usable medical devices as healthcare (clinical) waste in accordance with local healthcare guidelines, country and government legislation or policy.

Separation, re-cycling or disposal of packaging material shall follow local country and government legislation on packaging and packaging waste, where applicable.

Manufacturer and Distributor Information:



Manufacturer:

Nobel Biocare AB
Box 5190, 402 26
Västra Hamngatan 1
411 17 Göteborg
Sweden

www.nobelbiocare.com

Distributed in Australia by:

Nobel Biocare Australia Pty Ltd
Level 4/7 Eden Park Drive
Macquarie Park, NSW 2114 Australia
Phone: +61 1800 804 597

Distributed in New Zealand by:

Nobel Biocare New Zealand Ltd
33 Spartan Road
Takanini, Auckland, 2105 New Zealand
Phone: +64 0800 441 657



Note Regarding Canadian Device Licensure: Note that not all products described in this IFU may have been licensed in accordance with Canadian law.


Basic UDI-DI Information:

The following table lists the Basic UDI-DI information for the devices described in this IFU.

Product	Basic UDI-DI Number
Abutment Retrieval Instrument Zirconia CC	73327470000001747C
Abutment Retrieval Instrument Titanium CC	
Abutment Retrieval Tool Nobel Biocare N1™ TCC	

Symbols Glossary:

The following symbols may be present on the device labeling or in information accompanying the device. Refer to the device labeling or accompanying information for the applicable symbols.



Authorized representative in the European Community



Batch code



Catalogue number



Caution



CE marking



Consult instructions for use



Contains hazardous substances



Contains or presence of phthalate



Date



Date of manufacture



Do not re-sterilize



Do not re-use



Do not use if package is damaged



Double sterile barrier system

Rx Only

For prescription use only



Health care centre or doctor



Keep away from sunlight



Keep dry

symbol.glossary.nobelbiocare.com
ifu.nobelbiocare.com

Link to Online Symbols Glossary and IFU Portal



Magnetic resonance conditional



Manufacturer



Medical device



Non-pyrogenic



Non-sterile



Patient identification



Patient information website



Patient number



Serial number



Single sterile barrier system



Single sterile barrier system with protective packaging inside



Single sterile barrier system with protective packaging outside



Sterilized using ethylene oxide



Sterilized using irradiation



Temperature limit



Tooth number



Upper limit of temperature



Sterilized using steam or dry heat



Unique Device Identifier



Use-by date

EN All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images in this folder are not necessarily to scale. All product images are for illustration purposes only and may not be an exact representation of the product.