

On1 IOS (Intraoral Scannable) Healing Cap

Instructions for use



Important: Please read.

Disclaimer of liability:

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, it is his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof. Please note that some products detailed in this Instructions for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description:

The On1 IOS (Intraoral Scannable) Healing Cap is a prefabricated, implant supported component manufactured from radio opaque PEEK material.

The On1 IOS Healing Cap is available on Nobel Biocare's On1 Base, and comes in different platforms sizes and heights depending on the clinical situation.

The On1 IOS Healing Cap is delivered with a handle (to facilitate the insertion and mounting of the On1 IOS Healing Cap onto the On1 Base intraorally) and a Unigrip™ clinical prosthetic screw (to fasten the On1 IOS Healing Cap onto the On1 Base).

Intended use:

The On1 IOS Healing Cap is intended to be used in the upper or lower jaw to facilitate in sculpturing of soft tissue, during the healing phase post dental implant surgery.

The On1 IOS Healing Cap is intended to be used in the upper or lower jaw to facilitate digital capturing of the intraoral situation using an approved Nobel Biocare intraoral scanner, so that an individualized CAD/CAM restoration can be designed and ordered electronically, using a Nobel Biocare approved CAD system, and the individualized restoration fabricated in the dental laboratory, clinic or hospital.

The On1 IOS Healing Cap handle is intended to be used to facilitate the insertion and mounting of the On1 IOS Healing Cap onto the On1 Base intraorally.

The clinical prosthetic screw is intended to fasten the On1 IOS Healing Cap onto the On1 Base.

Indications for use:

The On1 IOS Healing Cap is a pre-manufactured adjustable prosthetic component to be directly connected to the On1 Base, which is fixated to an endosseous implant.

The On1 IOS Healing Cap is indicated for temporary use up to 180 days in prosthetic rehabilitation.

The On1 IOS Healing Cap is indicated for single-unit restorations.

Contraindications:

It is contraindicated to use the On1 IOS Healing Cap in:

- Patients who are medically unfit for an oral surgical procedure.
- Patients with inadequate bone volume unless an augmentation procedure can be considered.
- Patients in whom adequate sizes, numbers or desirable positions of implants are not reachable to achieve safe support of functional or eventually parafunctional loads.
- Patients who are allergic or hypersensitive to titanium alloy Ti-6Al-4V (titanium, aluminum, vanadium) and Polyetheretherketone (PEEK).

Warnings:

To secure the required functional accuracy for an intraoral CAD/CAM workflow, the On1 IOS Healing Cap should not be modified. If any modifications are made, the scanning function is impaired or removed.

Cautions:

Close cooperation between surgeon, restorative dentist and dental laboratory technician is essential for a successful implant treatment.

It is strongly recommended that the On1 IOS Healing Cap is used only with Nobel Biocare surgical instruments and prosthetic components, as combining components that are not dimensioned for correct mating can lead to mechanical and/or instrumental failure, damage to tissue or unsatisfactory esthetic results.

It is strongly recommended that clinicians, new as well as experienced implant users, always go through special training before undertaking a new treatment method. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more info please visit www.nobelbiocare.com.

Working the first time with a colleague, experienced with the new device/treatment method, avoids eventual complications. Nobel Biocare has a global network of mentors available for this purpose.

Special attention has to be given to patients who have local or systemic factors that could interfere with the healing process of either bone or soft tissue or the osseointegration process (e.g., cigarette smoking, poor oral hygiene, uncontrolled diabetes, oro-facial radiotherapy, steroid therapy, infections in the neighboring bone). Special caution is advised in patients who receive bisphosphonate therapy.

In general, implant placement and prosthetic design must accommodate individual patient conditions. In case of bruxism or unfavorable jaw relationships reappraisal of the treatment option may be considered.

With respect to pediatric patients, routine treatment is not recommended until the end of the jaw bone growth phase has been properly documented.

Pre-operative hard tissue or soft tissue deficits may yield a compromised esthetic result or unfavorable implant angulations.

All instruments and tooling used during procedure must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components.

Care and maintenance of instruments are crucial for a successful treatment. Sterilized instruments not only safeguard your patients and staff against infection but are also essential for the outcome of the total treatment.

Because of the small sizes of the devices, care must be taken that they are not swallowed or aspirated by the patient.

After the implant installation, the surgeon's evaluation of bone quality and primary stability will determine when implants may be loaded. Lack of adequate quantity and/or quality of remaining bone, infection and generalized diseases may be potential causes for failure of osseointegration both immediately after surgery, or after osseointegration is initially achieved.

To secure the long term treatment outcome, it is advised to provide comprehensive regular patient follow up after implant treatment and to inform about appropriate oral hygiene.

Handling procedures:

Clinical procedure:

- After placement of the On1 Base (according to the On1 Base Instructions for Use), for the healing phase and to facilitate the shaping of the soft tissue, insert and mount the On1 IOS Healing Cap on the On1 Base, using the provided handle (see Fig. 1) and carefully hand tighten the clinical prosthetic screw using an Unigrip™ screwdriver.
- The height of the On1 IOS Healing Cap should be chosen in accordance to the clinical situation (see Table A below). Additionally if a digital impression is going to be taken, see the Instructions for Use of the Intraoral Scanner manufacturer.

1



- **Note:** before inserting the On1 IOS Healing Cap onto the On1 Base ensure that the interface of the On1 Base is clean and free of foreign material that may obstruct the seating of the On1 IOS Healing Cap.

- Either a conventional or digital impression can be taken, following the steps defined below:

Conventional impression taking:

- Unfasten the clinical prosthetic screw using an Unigrip™ screwdriver and carefully remove the On1 IOS Healing Cap.

- Take a conventional impression according to the On1 Base Instructions for Use.

- After taking a conventional impression, carefully re-insert the On1 IOS Healing Cap onto the On1 Base and hand tighten the clinical prosthetic screw using an Unigrip™ screwdriver.

- Send the impression to the laboratory.

Digital impression taking:

- Take a digital impression of the On1 IOS Healing Cap, following the Intraoral scanner manufacturer's guidelines.

- Send the digital impression to the laboratory.

Caution: Never exceed **15 Ncm** torque for the On1 IOS Healing Cap clinical prosthetic screw. Overtightening of the clinical prosthetic screw may lead to a distortion of the On1 IOS Healing Cap.

The On1 IOS Healing Cap is available in different heights (see Table A).

A

Article Name	Height (mm)
On1 IOS Healing Cap 4.5mm NP	4.5
On1 IOS Healing Cap 6mm NP	6.0
On1 IOS Healing Cap 4.5mm RP	4.5
On1 IOS Healing Cap 6mm RP	6.0
On1 IOS Healing Cap 4mm WP	4.0
On1 IOS Healing Cap 5mm WP	5.0

Laboratory procedure:

– For conventional laboratory workflow procedures using an On1 Base, refer to the On1 Base Instructions for Use.

For procedures based on an intraoral scan, follow the below steps:

- Import the clinical situation into a Nobel Biocare approved CAD software
- Once imported, open the relevant CAD module and design your restoration, following the instructions in the software tutorial, according to the patient’s clinical needs.
- Send designed data file to your in-lab milling machine facility by clicking on the order button in the software, following the instructions in the software tutorial.
- Once the final restoration has been milled, finalize it according to standard laboratory procedures or material manufacturer’s instructions.
- Thereafter clean in ultrasonic unit.
- Send restoration to the clinician.

Clinical procedure:

- Unfasten the clinical prosthetic screw using an Unigrip™ screwdriver and carefully remove the On1 IOS Healing Cap.
- After removal of the On1 IOS Healing Cap refer to procedures defined in On1 Base Instructions for Use, for placement and tightening of the final restoration.

Materials:

On1 IOS Healing Cap: Polyetheretherketone (PEEK).

On1 IOS Healing Cap Handle: Polyetheretherketone (PEEK).

Clinical prosthetic screw: Titanium alloy 90% Ti, 6% Al, 4%V.

Cleaning and sterilization instructions:

The On1 IOS Healing Cap including handle and screw are delivered sterile and for single use only prior to the labelled expiration date.

Caution: The On1 IOS Healing Cap is a single use product and must not be reprocessed. Reprocessing could cause loss of mechanical, chemical and / or biological characteristics. Re-use could cause cross contamination.

Warning: Do not use device if the packaging has been damaged or previously opened.

Magnetic Resonance (MR) safety information:

The device has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of the device in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

For additional information on Cleaning and Sterilization and Magnetic Resonance Imaging, please consult the “Cleaning & Sterilization Guidelines for Nobel Biocare Products including MRI Information” available at www.nobelbiocare.com/sterilization or request latest printed version from a Nobel Biocare representative.

Storage and handling:

The product must be stored in a dry place in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage may influence device characteristics leading to failure.

Disposal:

Disposal of the device shall follow local regulations and environmental requirements, taking different contamination levels into account.



Manufacturer: Nobel Biocare AB, Box 5190, 402 26

Västra Hamngatan 1, 411 17 Göteborg, Sweden.

Phone: +46 31 81 88 00. Fax: +46 31 16 31 52. www.nobelbiocare.com

Canada – License Exemption: Please note that not all products may have been licensed in accordance with Canadian law.

Prescription Device – Rx only

Caution: Federal law restricts this device to sale by or on the order of a licensed physician or dentist.

Symbols Glossary:

The following symbols may be present on the device labeling or in information accompanying the device. Refer to the device labeling or accompanying information for the applicable symbols.



Authorized representative in the European Community



Batch code



Catalogue number



Caution



CE marking



Consult instructions for use



Contains hazardous substances



Contains or presence of phthalate



Date



Date of manufacture



Do not re-sterilize



Do not re-use



Do not use if package is damaged



Double sterile barrier system

Rx Only

For prescription use only



Health care centre or doctor



Keep away from sunlight



Keep dry



symbol.glossary.nobelbiocare.com
ifu.nobelbiocare.com

Link to Online Symbols Glossary and IFU Portal



Magnetic resonance conditional



Manufacturer



Medical device



Non-pyrogenic



Non-sterile



Patient identification



Patient information website



Patient number



Serial number



Single sterile barrier system



Single sterile barrier system with protective packaging inside



Single sterile barrier system with protective packaging outside



Sterilized using ethylene oxide



Sterilized using irradiation



Temperature limit



Tooth number



Upper limit of temperature



Sterilized using steam or dry heat



Unique Device Identifier



Use-by date

EN All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images in this folder are not necessarily to scale. All product images are for illustration purposes only and may not be an exact representation of the product.