

Multi-unit Aligning Instrument Instructions for use



Important: Please read.

Disclaimer of liability:

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, it is his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof. Please note that some products detailed in this Instructions for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description:

The Multi-unit Aligning Instrument is a reusable instrument to be attached to Nobel Biocare implant drivers. It has an angulation indicator to help the user to identify the angulation of the most suitable Multi-unit Abutment as well as the implant rotational position, that defines the abutment screw access hole.

Intended use:

The Multi-unit Aligning Instrument is a reusable instrument and it is intended to be used to identify the angulation of the best suitable Multi-unit Abutment as well as the implant rotational position, that defines the abutment screw access hole.

Indications:

The Multi-unit Aligning Instrument is indicated to be used as a reusable instrument during dental implant treatments with Nobel Biocare implants and restorative components.

Contraindications:

It is contraindicated using the Multi-unit Aligning Instrument in:

- Patients who are medically unfit for an oral surgical procedure.
- Patients who are contraindicated for treatment with Nobel Biocare implants or restorative components.
- Patients who are allergic or hypersensitive to stainless steel.

Cautions:

Close cooperation between surgeon, restorative dentist and dental laboratory technician is essential for a successful implant treatment.

It is strongly recommended that Nobel Biocare surgical instruments and prosthetic components are used only with Nobel Biocare implants, as combining components that are not dimensioned for correct mating can lead to mechanical and/or instrumental failure, damage to tissue or unsatisfactory esthetic results.

It is strongly recommended that clinicians, new as well as experienced implant users, always go through special training before undertaking a new treatment method. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more information, please visit www.nobelbiocare.com.

When using a new device/treatment method for the first time, working with a colleague who is experienced with the new device/treatment method may help avoid possible complications. Nobel Biocare has a global network of mentors available for this purpose.

All instruments and tooling used in surgery must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components. Care and maintenance of instruments are crucial for a successful treatment. Sterilized instruments not only safeguard your patients and staff against infection but are also essential for the outcome of the total treatment.

Because of the small size of the devices, care must be taken that they are not swallowed or aspirated by the patient.

Handling procedure:

Multi-unit Aligning Instrument handling procedure:

Please place the appropriate implant following corresponding Instructions For Use. Once the implant has been inserted and seated, proceed as follows:

1. Secure the Multi-unit Aligning Instrument with dental floss as shown in **Figure 1a**.
2. Assemble the Multi-unit Aligning Instrument on the implant driver see **Figure 1a**. If a Tri-Channel implant driver is used, then the laser marking on the implant driver (red arrow) has to be aligned with the Multi-unit Aligning Instrument as shown in **Figure 1b**.

Figure 1a



Figure 1b



3. Insert the implant driver Multi-unit Aligning Instrument assembly into the implant see **Figure 2**.

Figure 2



4. The angulation indicator of the Multi-unit Aligning Instrument indicates the position of the prosthetic screw hole when selecting a 17° or 30° Multi-unit Abutment. Position the Multi-unit Aligning Instrument so that the angulation indicator is perpendicular to the bone. The arm of the Multi-unit Aligning Instrument perpendicular to the bone, indicates the recommended Multi-unit Abutment to use. The red line in **Figure 3** illustrates the 30° option.

Figure 3



5. Adjust the rotational position of the implant if necessary with the Manual Torque Wrench see **Figure 4**.

Figure 4



Caution: Never exceed insertion torque detailed in the respective implant Instructions For Use. Overtightening of implant may lead to damage of the implant, fracture or necrosis of the bone site. If a Surgical Driver is used to insert the implant, special care needs to be taken to avoid over tightening.

6. Remove the implant driver and Multi-unit Aligning Instrument assembly and insert the appropriate Multi-unit Abutment.

Figure 5 illustrates a 30° Multi-unit Abutment including the insertion handle.

Figure 5



For additional information on surgical and restorative procedures please consult treatment guidelines available at www.nobelbiocare.com or request latest printed version from a Nobel Biocare representative.

Material:

The Multi-unit Aligning Instrument is made of stainless steel.

Cleaning and sterilization instructions:

The Multi-unit Aligning Instrument is delivered non-sterile and intended for re-use. Prior to use and re-use clean, disinfect and sterilize the product using the recommended parameters.

Warning: Do not use device if the packaging has been damaged or previously opened.

Warning: Use of non-sterile components may lead to infection of tissues or infectious diseases.

Multi-unit Aligning Instrument must be disassembled from the implant driver prior to cleaning.

Clean the device using manual or automated cleaning, disinfect and dry the device following instructions in Cleaning and Sterilization Guidelines available at <http://www.nobelbiocare.com/sterilization>.

Inspect and seal the single device in a pouch and steam sterilize, both the gravity cycle (saturated steam) and pre-vacuum (forced air removal) cycle can be applied, using the following parameters:

For USA: Steam sterilization 270°F (132°C) for 4 minutes when using pre-vacuum method and 15 minutes when using the gravity method. Dry for 20 to 30 minutes when using pre-vacuum method and 15 to 30 minutes when using the gravity method.

For USA: FDA-cleared sterilization accessories are to be used for the recommended sterilization parameters.

For outside USA: Temperature 132°C (270°F), max 137°C (279°F) for 3 minutes (up to 20 minutes). Dry for 10 minutes in chamber.

Alternative UK: Temperature 134°C (273°F), max 137°C (279°F) for 3 minutes (up to 20 minutes). Dry for 10 minutes in chamber.

Full set of recommended parameters are provided in "Cleaning & Sterilization Guidelines including Magnetic Resonance Imaging Information of Nobel Biocare Products" available at www.nobelbiocare.com/sterilization or request latest printed version from a Nobel Biocare representative.

Storage, handling and transportation:

The device must be stored and transported in dry conditions in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage and transportation may influence device characteristics leading to failure.

Disposal:

Disposal of the device shall follow local regulations and environmental requirements, taking different contamination levels into account.



Manufacturer: Nobel Biocare AB, Box 5190, 402 26
Västra Hamngatan 1, 411 17 Göteborg, Sweden.

Phone: +46 31 81 88 00. Fax: +46 31 16 31 52. www.nobelbiocare.com

Canada license exemption: Please note that not all products may have been licensed in accordance with Canadian law.

For Prescription Use Only. Caution: Federal (United States) law restricts this device to sale by or on the order of a clinician, medical professional or physician.



Rx Only



Non-sterile



Caution



Consult instructions
for use



Batch code



Do not use if package
is damaged

EN All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images are not necessarily to scale.