

Tapered Drills and Screw Taps – Reusable

Instructions for use



Important: Please read.

Disclaimer of liability:

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, they are his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof. Please note that some products detailed in this Instruction for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description:

Reusable Tapered Drills, Dense Bone Drills and Screw Taps are made of stainless steel with DLC (Diamond Like Carbon) coating and should be replaced after 20–30 uses, or when cutting efficiency declines. Tapered Drills are internally-irrigated and require a specific technique to prevent irrigation holes becoming plugged with bone. Tapered drills are unique for each implant length.

Intended use:

Tapered Drills, Dense Bone Drills and Screw Taps are to be used in the upper or lower jaw to prepare osteotomy prior to implant placement.

Indications:

Tapered Drills, Dense Bone Drills and Screw Taps are to be used in combination with NobelReplace® Tapered Groovy®, Replace Select™ Tapered TiUnite®, Replace Select™ Tapered Partially Machined Collar (PMC), NobelReplace® Conical Connection (CC), NobelReplace® Conical Connection Partially Machined Collar (CC PMC), NobelDirect® Oval, and NobelDirect® Posterior implants.

Contraindications:

In general, contraindications are applicable for implant surgery related procedures in patients:

- who are medically unfit for an oral surgical procedure.
- who are allergic or hypersensitive to stainless steel or diamond-like carbon (DLC) coating.
- in whom adequate sizes, numbers or desirable positions of implants are not reachable to achieve safe support of functional or eventually parafunctional loads.

Warnings:

Failure to recognize actual lengths of drills relative to radiographic measurements can result in permanent injury to nerves or other vital structures. Drilling beyond the depth intended from lower jaw surgery may potentially result in permanent numbness to the lower lip and chin or lead to a hemorrhage in the floor of the mouth.

Besides the mandatory precautions for any surgery such as sepsis, during drilling in the jaw bone, one must avoid damage the nerves and vessels by referring to anatomical knowledge and preoperative radiographs.

Cautions:

General:

It is strongly recommended that surgical instruments are used only with Nobel Biocare implants as detailed, as combining components that are not dimensioned for correct mating can lead to mechanical and/or instrumental failure, damage to tissue or unsatisfactory esthetic results.

It is strongly recommended that clinicians, new as well as experienced implant users, always go through special training before undertaking a new treatment method. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more info please visit www.nobelbiocare.com.

Working the first time with a colleague, experienced with the new device/treatment method, avoids eventual complications. Nobel Biocare has a global network of mentors available for this purpose.

Before surgery:

Careful clinical and radiological examination of the patient has to be performed prior to surgery to determine the psychological and physical status of the patient.

Pre-operative hard tissue or soft tissue deficits may yield a compromised esthetic result or unfavorable implant angulation.

At surgery:

All instruments and tooling used during procedure must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components.

Because of the small size of components, care must be taken that they are not swallowed or aspirated by the patient.

After surgery:

To secure the long term treatment outcome it is advised to provide comprehensive regular patient follow up after implant treatment and to inform about appropriate oral hygiene.

Surgical procedures:

Drilling must proceed at high speed for Tapered Drills (maximum 800rpm) under constant and profuse irrigation by sterile saline at room temperature. Tapered Drills are internally-irrigated and require a specific technique to prevent irrigation holes becoming plugged with bone debris. During drilling use an in-and-out motion and drill in bone for 1–2 seconds. Move the drill up without stopping handpiece motor which allows the irrigation to flush away bone debris.

Caution: Tapered Drills extend up to 1 mm longer than the implant when seated.

Allow for this additional length when drilling near vital anatomical structures.

Dense bone protocol (optional) – as indicated:

Dense Bone Drill is only needed for 13mm and 16mm implants. If shorter implants are used, go directly to step (b).

- Select the Dense Bone Drill matching the diameter and length (13 or 16mm) of final Tapered Drill. Drill one pass into the prepared site with high speed (800rpm) using Bone Drill.
- Select the Screw Tap matching the diameter of final Tapered Drill. Place into prepared implant site using low speed (25rpm).

- Apply firm pressure and begin rotating the Screw Tap slowly. When the threads engage, allow Screw Tap to feed without pressure to appropriate depth.
- Switch the handpiece to reverse mode and back the Screw Tap out.

For additional information on surgical procedures please consult the NobelReplace® Tapered Groovy®, Replace Select™ Tapered TiUnite®, Replace Select™ Tapered Partially Machined Collar (PMC), NobelReplace® Conical Connection (CC), NobelReplace® Conical Connection Partially Machined Collar (CC PMC), NobelDirect® Groovy®, NobelDirect® Oval, and NobelDirect® Posterior “Procedures & products” treatment guidelines available at www.nobelbiocare.com or request latest printed version from a Nobel Biocare representative.

Materials:

Tapered Drills, Dense Bone Drills and Screw Taps: stainless steel, DLC (Diamond Like Carbon) coating.

Cleaning and sterilization:

Tapered Drills, Dense Bone Drills and Screw Taps are delivered non-sterile and must be cleaned and sterilized prior to use.

For USA: Seal single device in a pouch and steam sterilize at 270° F (132° C) for 3 minutes.

For outside USA: Seal single device in a pouch and steam sterilize at 132° C–135° C (270° F–275° F) for 3 minutes.

Alternative UK: Seal single device in a pouch and steam sterilize at 134° C–135° C (273° F–275° F) for 3 minutes.

Warning: Use of non-sterile components may lead to infection of tissues or infectious diseases.

Full set of recommended parameters are provided in “Cleaning & Sterilization Guidelines including MRI Information of Nobel Biocare Products” available at www.nobelbiocare.com/sterilization or request latest printed version from a Nobel Biocare representative.

MRI safety information:

Please note that the product has not been evaluated for safety and compatibility in the MR environment. The product has not been tested for heating or migration in the MR environment.

For additional information on Magnetic Resonance Imaging, please consult the “Cleaning & Sterilization Guidelines for Nobel Biocare Products including MRI Information” available at www.nobelbiocare.com/sterilization or request latest printed version from a Nobel Biocare representative.

Storage and handling:

The product must be stored in a dry place in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage may influence device characteristics leading to failure.

Disposal:

Disposal of the device shall follow local regulations and environmental requirements, taking different contamination levels into account.

Manufacturer: Nobel Biocare AB, Box 5190, 402 26

Västra Hamngatan 1, 411 17 Göteborg, Sweden.

Phone: +46 31 81 88 00. Fax: +46 31 16 31 52. www.nobelbiocare.com

Canada license exemption: Please note that not all products may have been licensed in accordance with Canadian law.

Prescription device: Rx only

Caution: Federal law restricts this device to sale by or on the order of a licensed physician or dentist.

CE 0086

Rx Only



Non-sterile



Consult instructions
for use



Batch code

US All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images in this folder are not necessarily to scale.