

Position Locator



Important – Disclaimer of Liability

This product is part of an overall concept and may only be used in conjunction with the associated original products according to the instructions and recommendation of Nobel Biocare. Non-recommended use of products made by third parties in conjunction with Nobel Biocare products will void any warranty or other obligation, express or implied, of Nobel Biocare. The user of Nobel Biocare products has the duty to determine whether or not any product is suitable for the particular patient and circumstances. Nobel Biocare disclaims any liability, express or implied, and shall have no responsibility for any direct, indirect, punitive or other damages, arising out of or in connection with any errors in professional judgment or practice in the use of Nobel Biocare products. The user is also obliged to study the latest developments in regard to this Nobel Biocare product and its applications regularly. In cases of doubt, the user has to contact Nobel Biocare. Since the utilization of this product is under the control of the user, they are his/her responsibility. Nobel Biocare does not assume any liability whatsoever for damage arising thereof.

Please note that some products detailed in this Instruction for Use may not be regulatory cleared, released or licensed for sale in all markets.

Description

Position Locators are premanufactured dental prosthetic components which are connected to an endosseous dental implant/abutment placed in the patient's mouth, or to an implant/abutment replica embedded in a master cast, to facilitate the design and fabrication of a dental implant restoration. Position Locators Nobel Biocare N1™ TCC are available in NP/RP platforms, feature a tri-oval conical connection

and can be used with Nobel Biocare's Nobel Biocare N1™ implant system. A new version of Position Locators Nobel Biocare N1™ TCC has been introduced where the screw, which is used to attach the position locator to the implant or implant replica, is co-packed and can be assembled from the side window. Furthermore the screwdriver access has been added also on the upper side and the implant platform/connection information is now engraved on the side. The old version of the Position Locator Nobel Biocare N1™ TCC presents the screw preassembled and has the screwdriver access only through the side window. Nobel Biocare Multi-unit PoLo is available in two different heights and is compatible with the Multi-unit abutments of the corresponding platform. The two heights present a different matching surface and a line at 8mm for an immediate visual feedback of the dimension. The Nobel Biocare Multi-unit PoLo for Multi-unit abutments can be used together with Nobel Biocare Multi-Unit PoLo Link to enhance the scanning capability especially in edentulous cases. The links are available in three different lengths: 10 mm, 15 mm, 20 mm. Nobel Biocare Multi-unit PoLo is delivered copacked with the prosthetic screw used to attach the device to the Multi-unit abutment.

The following images represent the Nobel Biocare Multi-unit PoLo and Nobel Biocare Multi-unit PoLo Link.



Figure A – Nobel Biocare Multi-unit PoLo 8 mm and Nobel Biocare Multi-unit PoLo 11 mm



Figure B – Nobel Biocare Multi-unit PoLo Link



Figure C – Nobel Biocare Multi-unit PoLo assembled with Nobel Biocare Multi-unit PoLo Link



Figure D – Old (without engraving) and new (with engraving) Position Locator Nobel Biocare N1™ TCC

Nobel Biocare products are intended and available to be used in a variety of configurations. For further information refer to Nobel Biocare publication Compatibility Information by navigating to ifu.nobelbiocare.com.

Intended Use/Intended Purpose

Intended for use to transfer the direction, position or orientation of a dental implant to a patient model.

Indications

Position Locators Nobel Biocare N1™ TCC and Nobel Biocare Multi-unit Polo are indicated for use in combination with an intra-oral or desktop scanner to confirm the location, position and angulation of a dental implant/abutment or dental implant/abutment replica to support creation of the digital model to facilitate the design and fabrication of a dental prosthesis using CAD/CAM technology.

Contraindications

It is contraindicated to use the position locators and accessories in:

- Patients who are medically unfit for an oral surgical procedure.
- Patients in whom adequate sizes, numbers or desirable positions of implants are not reachable to achieve safe support of functional or eventually parafunctional loads.
- Patients who are allergic or hypersensitive to titanium alloy Ti-6Al-4V (titanium, aluminum, vanadium) or ZrN (zirconia nitride).

Materials

- Position Locator Nobel Biocare N1™ TCC, Nobel Biocare Multi-unit PoLo, and Nobel Biocare Multi-unit PoLo Link: Titanium alloy 90% Ti, 6% Al, 4% V according to ASTM F136 and ISO 5832-3, Zirconia nitride coating 58% Zr, 42% N.
- Screw: Titanium alloy 90% Ti, 6% Al, 4% V according to ASTM F136 and ISO 5832-3

Warnings

The devices shall be inspected before each re-use to ensure that the integrity and performance continues to be maintained. Inspect the devices for signs of degradation that may limit the useful life of the devices, such as:

Use of non-sterile device may lead to infection of tissues or infectious diseases.

Do not use device if the packaging has been damaged or previously opened as the device sterility and/or integrity may be compromised.

Cautions

General

Close cooperation between surgeon, restorative dentist and dental laboratory technician is essential for a successful implant treatment.

Nobel Biocare Multi-unit PoLo, Position Locators Nobel Biocare N1™ TCC and Nobel Biocare Multi-unit PoLo Link must only be used with compatible Nobel Biocare instruments and/or components and/or prosthetic components. Use of instruments and/or components and/or prosthetic components that are not intended to be used in combination with Nobel Biocare Multi-unit PoLo, Position Locators Nobel Biocare N1™ TCC and Nobel Biocare Multi-unit PoLo Link can lead to product failure, damage to tissue, or unsatisfactory esthetic results.

When using a new device/treatment method for the first time, working with a colleague who is experienced with the new device/treatment method may help avoid possible complications. Nobel Biocare has a global network of mentors available for this purpose.

Before Surgery

Careful psychological and physiological evaluation, followed by clinical and radiological examination must be performed on the patient prior to surgery to determine the suitability of the patient for treatment.

Special attention must be given to patients who have local or systemic factors that could interfere with the healing process of either bone or soft tissue or the osseointegration process (e.g. cigarette smoking, poor oral hygiene, uncontrolled diabetes, oro-facial radiotherapy, steroid therapy, infections in the neighboring bone). Special caution is advised in patients who receive bisphosphonate therapy.

In general, implant placement and prosthetic design must accommodate individual patient conditions. In case of bruxism, other parafunctional habits or unfavorable jaw relationships, reappraisal of the treatment option may be considered.

The device has not been evaluated in pediatric/adolescent patients and is not recommended for use in children. Routine treatment is not recommended until the end of the juvenile jaw bone growth phase has been properly documented.

Pre-operative hard tissue or soft tissue deficits may yield a compromised esthetic result or unfavorable implant angulations.

All components, instruments and tooling used during the clinical and/or laboratory procedure must be maintained in good condition and care must be taken that instrumentation does not damage implants or other components.

At Surgery

Care and maintenance of sterile instruments are crucial for a successful treatment. Sterilized instruments not only safeguard your patients and staff against infection but are also essential for the outcome of the total treatment.

Because of the small sizes of the devices, care must be taken that they are not swallowed or aspirated by the patient. It is appropriate to use specific supporting tools to prevent aspiration of loose parts (e.g. gauze, dental dam, or throat shield).

After Surgery

To help ensure a successful long term-treatment outcome, it is advised to provide comprehensive regular patient follow up after implant treatment and to inform the patient about appropriate oral hygiene.

Intended Users and Patient Groups

Position Locator Nobel Biocare N1™ TCC, Nobel Biocare Multi-unit PoLo, Nobel Biocare Multi-unit PoLo Link are to be used by dental health care professionals.

Position Locator Nobel Biocare N1™ TCC, Nobel Biocare Multi-unit PoLo, Nobel Biocare Multi-unit PoLo Link are to be used in patients subject to dental implant treatment.

Clinical Benefits and Undesirable Side Effects

Clinical Benefits Associated with Devices in the IFU

Position Locator Nobel Biocare N1™ TCC, Nobel Biocare Multi-unit PoLo, Nobel Biocare Multi-unit PoLo Link are components of treatment with a dental implant system and/or dental crowns and bridges. As a clinical benefit of treatment, patients can expect to have their missing teeth replaced and/or crowns restored.

Undesirable Side Effects Associated with Position Locator Nobel Biocare N1™ TCC, Nobel Biocare Multi-unit PoLo, Nobel Biocare Multi-unit PoLo Link:

The placement of these devices is part of an invasive treatment which may be associated with typical side effects such as inflammation, infection, bleeding, hematoma, pain, swelling. Depending on location it may also lead in rare cases to fenestration or fracture of bone, damage/perforation of neighboring structures/restorations, sinusitis, or sensory/motor disturbances. During use of these devices the pharyngeal (gag) reflex may be triggered in patients with a sensitive gag reflex.

Where required per the European Medical Device Regulation (MDR; EU 2017/745) a Summary of Safety and Clinical Performance document (SSCP) is available for the Position Locator. The SSCP can be obtained at the following website: ec.europa.eu/tools/eudamed¹

¹ Website available upon launch of the European Database on Medical Devices (EUDAMED)

Notice regarding serious incidents

For a patient/user/third party in the European Union and in countries with an identical regulatory regime (Regulation 2017/745/EU on Medical Devices); if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and to your national authority. The contact information for the manufacturer of this device to report a serious incident is as follows:

Nobel Biocare AB
www.nobelbiocare.com/complaint-form

Handling Procedure

Note Before each use inspect the devices for scratches and deformation to ensure that the integrity and performance of the position locator and link are maintained.

Position Locator Nobel Biocare N1™ TCC Intra-oral Scan Workflow

Assemble the new version of the position locator:

1. Insert the screw from the side slot into the screw hole.



Figure E – Screw assembled through the side window of the new Nobel Biocare N1™ TCC Position Locator

2. Insert the Omnigrip™ Mini screwdriver through the center hole.
3. Engage the screw head and tight the screw into the Position Locator body.
4. Once you see the entire screw thread, the screw is fully inserted.



Figure F – Insertion and screw tightening from the top of Omnigrip™ Mini Screwdriver for the new version of Position Locator Nobel Biocare N1™ TCC

Handling procedure for the old and new version of Position Locator Nobel Biocare N1™ TCC:

- Connect the Position Locator to the implant by hand-tightening the screw using the Omnigrip™ Mini screwdriver. For the new version of the position locator, it is recommended to use the top hole for the screw tightening whenever the space allows it, as then the screwdriver and the screw are completely aligned. For the old version of the position locator use the side access. Refer to Nobel Biocare Instructions for Use (IFU) IFU1085 for information regarding the screwdrivers.



Figure G – Tightening from the top for the new version and from the side window for the old version of Position Locator Nobel Biocare N1™ TCC

- Verify the seating of the position locator using radiographic imaging.
- Take an intra-oral scan of the patient following the scanner manufacturer's instructions. The new position locator present the type of connection and implant platform engraved on the side which can be captured by the scanner

Engraving	Implant connection	Platform
N1R	(N1) for Nobel Biocare N1™ TCC	(R) for RP
N1N	(N1) for Nobel Biocare N1™ TCC	(N) for NP



Figure H – Example of engraving on the side of the new version of the Position Locator Nobel Biocare N1™ TCC

- Remove the Position Locator by untightening the screw.
- Send the scan file to the laboratory.

Position Locator Nobel Biocare N1™ TCC Desktop Scanning Workflow

- Connect and hand-tighten the Position Locator by tightening the screw to the implant replica embedded in the master cast using the Omnigrip™ Mini screwdriver. Refer to Nobel Biocare Instructions for Use (IFU) IFU1085 for information regarding the screwdrivers.
- Scan the master cast following the scanner manufacturer's instructions.

Nobel Biocare Multi-unit PoLo Intra-oral Scan Workflow

- Select the Appropriate position locator height and link length:
 - Evaluate the specific requirements of the case to determine the ideal position locator height and links length.
 - The links are recommended to be added for edentulous cases and are available in various lengths to accommodate different implant positions.
- If needed, assemble the links to the position locators:
 - Perform this step extra-orally.
 - Manually tighten the links to the position locators and insert the copacked prosthetic screw into the position locators using the Unigrip™ screwdriver.
- Placement of Nobel Biocare Multi-unit PoLo on Multi-Unit Abutments:
 - Place the position locators on the Multi-unit abutments.
 - Partially tighten them to ensure stability and align the links toward the next position locator, following the natural arch shape.

For full arch cases, align the links to follow the natural arch shape and position it close to the next position locator. Refer to the provided images for an example setup involving four implants.



Figure I – Example of All-on-4® configuration

- Hand-Tighten the Screw:
 - Once each position locator and link are aligned, use the Unigrip™ screwdriver to securely fasten the screw.

Caution Avoid excessive force on the link as it may damage the Multi-unit abutment/implant connection.

Note Ensure that the link does not exert pressure on the surrounding soft tissue. A minor non-parallel alignment between the links and the tissues, will not compromise the scan.

5. Intra-Oral Scanning technique:
 - Begin scanning from the posterior position locator. Keep the scanner tip occlusal during scanning.
 - Proceed with a continuous scan from one direction to the other. Aim to capture as much information as possible using linear movements. Slightly tilt the tip of the scanner in order to capture the top of the position locator and the soft tissue paying attention to remove potential scanning gaps between the position locator and the tissue.
 - Once the first scan is completed, if gaps are detected, scan those missing areas individually to ensure comprehensive coverage.

Note Be cautious not to touch the Links during scanning, as any movement could compromise the final result.

6. Save Digital Scan Data:
 - Save the digital scan data files and send the files to the laboratory.
7. Removing Position Locators:
 - Carefully untighten the position locators using the Unigrip™ screwdriver

Additional consideration:

- For a complete profile of the soft tissue, perform a full arch scan without the position locators in order to capture the whole profile of the soft tissue.
- Before proceeding with the final restoration, consider testing the passive fit of a prototype.

Nobel Biocare Multi-unit PoLo Desktop Scanning Workflow

1. Connect and hand-tighten the position locator by tightening the screw to the implant replica embedded in the master cast using the Unigrip™ screwdriver. Refer to Nobel Biocare Instructions for Use (IFU) IFU1085 for information regarding the screwdrivers.
2. Scan the master cast following the scanner manufacturer's instructions.

Sterility and Reusability Information

Nobel Biocare Position Locators are delivered non-sterile and are intended for reuse. Prior to use clean and sterilize the product following the manual or automated procedure in the Cleaning and Sterilization Instructions.

Warning Use of non-sterile device may lead to infection of tissues or infectious diseases.

Warning Do not use device if the packaging has been damaged.

The devices shall be inspected before each re-use to ensure that the integrity and performance continues to be maintained. Inspect the devices for signs of degradation that may limit the useful life of the devices, such as:

- Any sign of wear or modification of the surface.
- Visible corrosion.
- Mechanical wear/damage.

The device shall be disposed if any of these signs of degradation are evident.

Cleaning and Sterilization Instructions

These products are intended to be cleaned and sterilized. For further information refer to Nobel Biocare publication **Cleaning and Sterilization Instructions** by navigating to ifu.nobelbiocare.com.

Performance Requirements and Limitations

To achieve the desired performance, the devices must only be used with the products described in this Instructions for Use and/or in the Instructions for Use for other compatible Nobel Biocare products, and in accordance with the Intended Use for each product. To confirm the compatibility of products which are intended to be used in conjunction with the devices, check the color coding, dimensions, lengths, connection type and/or any direct marking as applicable on the products or product labeling.

Facilities and Training

It is strongly recommended that new and experienced users of Nobel Biocare products always go through special training before using a new product for the first time. Nobel Biocare offers a wide range of courses for various levels of knowledge and experience. For more information, please visit www.nobelbiocare.com.

Storage, Handling and Transportation





The device must be stored and transported in dry conditions in the original packaging at room temperature and not exposed to direct sunlight. Incorrect storage and transportation may influence device characteristics leading to failure.

Disposal

Safely discard potentially contaminated or no longer usable medical devices as healthcare (clinical) waste in accordance with local healthcare guidelines, country and government legislation or policy.

Separation, re-cycling or disposal of packaging material shall follow local country and government legislation on packaging and packaging waste, where applicable.

Manufacturer and Distributor Information

Manufacturer 	Nobel Biocare AB PO Box 5190, 402 26 Västra Hamngatan 1 Göteborg 411 17 Sweden www.nobelbiocare.com
UK Responsible Person <div>UK RP</div>	Nobel Biocare UK Ltd 4 Longwalk Road Stockley Park Uxbridge UB11 1FE United Kingdom
Distributed in Turkey by	EOT Dental Sağlık Ürünleri ve Dış Ticaret A.Ş. Nispetiye Mah. Aytar Cad. Metro İş Merkezi No: 10/7 Beşiktaş İSTANBUL Phone: +90 2123614901, Fax: +90 2123614904
Distributed in Australia by	Nobel Biocare Australia Pty Ltd Level 4, 7 Eden Park Drive Macquarie Park, NSW 2113 Australia Phone: +61 1800 804 597
Distributed in New Zealand by	Nobel Biocare New Zealand Ltd 33 Spartan Road Takanini, Auckland, 2105 New Zealand Phone: +64 0800 441 657
CE Mark for Class I Devices	
CE Mark for Class II Devices	
UKCA Mark for Class I Devices	

Note Refer to the product label to determine the applicable conformity marking for each device.

Note Regarding Canadian device licensure, not all products described in the IFU may have a device licence according to Canadian law.

Basic UDI-DI Information

Product	Basic UDI-DI Number
Position Locator Nobel Biocare N1™ TCC NP/RP	733274700000013674
Nobel Biocare Multi-unit PoLo	

Legal Statements

EN All rights reserved.

Nobel Biocare, the Nobel Biocare logotype and all other trademarks used in this document are, if nothing else is stated or is evident from the context in a certain case, trademarks of Nobel Biocare. Product images in this folder are not necessarily to scale. All product images are for illustration purposes only and may not be an exact representation of the product.

Symbols Glossary

Please refer to the packaging label for the applicable symbols related to the product. On the packaging label you may encounter various symbols to convey a specific information about the product and/or its use. For further information refer to Nobel Biocare publication to the **Symbols Glossary** by navigating to ifu.nobelbiocare.com.